

Materials List for:

Quantitative Autonomic Testing

Peter Novak¹


¹Department of Neurology, University of Massachusetts Medical School

Correspondence to: Peter Novak at novakp@umhc.org

URL: <https://www.jove.com/video/2502>

DOI: [doi:10.3791/2502](https://doi.org/10.3791/2502)

Materials

Name	Company	Catalog Number	Comments
Table 9: Recommended equipment.			
Motorized tilt table	WR Medical Electronics Co.		Any table with smooth movement reaching 70 degree within 5-10 seconds
ECG monitor			Any device with analog output
Noninvasive beat-to-beat blood pressure monitor	Finapres Medical Systems		Several models are available, all are similar
Q-Sweat	WR Medical Electronics Co.		
Transcranial Doppler			Any device with analog output
Recording system	 ADInstruments		Simple programming
Recording system	Labview		Complex programming
Recording system	Matlab		Complex programming
Recording system	WR Medical Electronics Co.		Ready to use, but limited flexibility
CO2 monitor			Many models used for sleep studies
Respiratory monitor			Simple thermistor attached to a recording system